

Walkthrough: Extraction of data from Exiqon arrays, using Feature Extraction

Before starting, make sure tiff image is in correct orientation (two landing lights in lower right corner).
If image was scanned using an Agilent scanner, the image must be flipped from upper left to lower right.

Protocol



If using extended dynamic range images (Agilent scanner), be sure that the files are named XXX_H.tif and XXX_L.tif